Active®

Best Design, best function

A new generation rollator

CE

ONE CONTENTS

1. Contents
2. Introduction
3. Parts Description
4. Personal Safety
5. Adjustments for Comfort
6. Transportation and Assembly
7. Care and Maintenance
8. Operation
9. Specification
10. Warranty

TWO INTRODUCTION

The Wheeltech Active rollator combines all of the best user-friendly features in a modern and sleek design. The rollator is designed to aid the user in a number of everyday tasks such as walking and shopping.

The rollator is suitable for everyday indoor and outdoor use. The rollator features extra-large front wheels which lets the rollator roll over obstacles and aids kerb climbing.

All Wheeltech rollators are manufactured to the highest quality standards and are CE marked as Medical Devices. The Active Rollator provides a robust and visually appealing product whilst being a cost-effective solution to some mobility problems.

THREE PARTS DESCRIPTION

FOUR PERSONAL SAFETY

- For your health and comfort, make sure that you adopt a comfortable posture and that you can operate the braking mechanism with ease.
- Please store the unit in a safe environment.
- Ensure the rollator is always fully folded out before using.
- Ensure that all the wheels are stable and are on the ground before using.
- Ensure the braking mechanism is adequate and test before using.
- Ensure that the user sits in the centre of the seat, with the seat at a height where the user has their feet on the floor when sat.
- Ensure that the brakes are always locked on when sitting on the seat.
- Ensure that the latest health and safety practice is used for carrying and lifting items, such as the Active Rollator.
- The rollator can only be used by one person at a time. Do not allow a user to be pushed whilst sat on the rollator.

FIVE ADJUSTMENTS FOR COMFORT

Handle Height Adjustment

To adjust the handle height of the rollator, push the Handle Height Adjuster out whilst repositioning the handle. Once the correct height has been selected release the Adjuster and pull the handle up until it locks in to position.

Repeat for the other handle.

Seat Height Adjustment

To adjust the height of the seat, pull both Seat Height Adjusters whilst repositioning the seat, then release both Adjusters and lock the seat in to position. See the photograph on the left.

SIX TRANSPORTATION & ASSEMBLY

- To fold up the rollator for transporting or storage, pull up the Folding Lever. The rollator should look like the photo (left)
- To unfold the rollator for use, push down on the Folding Lever.
- To gain access to the basket, lift up the seat. The basket can be removed for carrying shopping and other items.

SEVEN CARE AND MAINTENANCE

- Ensure the rollator is used only as an aid for mobility.
- Ensure the rollator is kept clean. The rollator can be wiped down with a cloth or sponge using mild detergent, or can be cleaned with a high pressure washer or hosepipe.
- Check the moving parts (e.g. wheels and height adjusters) regularly.
- Ensure that the handle operated cable brakes have been adjusted to suit the wheels and user.
- All the fasteners (screws, nuts and bolts) are to be checked regularly to ensure they are securely tightened.
- Only use an approved Medicare Technology dealer for any maintenance or repairs to your rollator.

EIGHT OPERATION

Basket.

To gain access to the basket, lift up the seat. The basket can be removed for carrying shopping and other items.

Obstacles.

The rollator comes with extra large front castors and a stepper pad to negotiate large obstacles (such as a kerb). Press down on the stepper pad with your foot to lift the front wheels to aid obstacle negotiation.

Brakes.

Squeeze the brake lever up to engage the brakes. To lock the brakes, push down on the brake lever until it locks. To unlock the brakes, pull up on the brake lever until the lock removes.

The brake tension can be adjusted by the bolt located underneath the stepper tube.

Using the allen key provided adjust the screw as shown in the photograph (left). If braking force is inadequate, screw in the bolt slightly to increase the tension. If the brakes are too difficult to apply, screw out the bolt slightly to reduce the tension.

Note: After adjusting the brake tension, check the that the brakes are comfortable to use and apply enough force to the wheels to stop the rollator.

NINE SPECIFICATION

The rollator comes with the following as standard:

- Aluminium frame and handlebars
- Height adjustable handles with lockable cable brakes
- Extra large front wheels and stepper tube
- Height adjustable seat with cupholders and collapsible & removable basket.
- Puncture-proof solid grey tyres

Technical data

Type	Active
User height	160 – 190 cm (5'3" – 6'3")
Max. user weight	125 kg (275 lbs, 19.8 stone)
Size in cm (wxlxh)	60 x 84 x 82–102
Size folded	60 x 50 x 89
Weight	< 9 kg (20 lbs)
Material	Aluminium and plastic
Cleaning	Tolerates high pressure wash

TEN WARRANTY

To improve the safety and lifespan of your rollator, it is recommended that you have it serviced on a regular basis. Recommended Service Schedules for rollators are available to service agents.

There is a comprehensive twelve-month warranty from the date on which your new rollator is delivered. The warranty covers the rollator for repairs or replacement during this period. For more detail, please see the warranty conditions below :

1. Any work or replacement part installation must be carried out by an authorised Medicare Technology dealer / service agent.
2. To apply the warranty should your rollator require attention please contact the outlet from which you purchased the rollator.
3. Should any part of the rollator require repair or full or part replacement, as a result of a manufacturing or material defect within the warranty period, parts will be supplied free of charge. Note: The guarantee is not transferable.
4. Any repaired or replaced parts will be covered by the balance of the warranty period on the rollator.
5. Parts replaced after the original warranty has expired will be covered by a three-month warranty.
6. Consumable items supplied will not generally be covered during the normal warranty period unless such items require repair or replacement clearly as a direct result of a manufacturing or material defect. Such items include (among others): upholstery and tyres.
7. The above warranty conditions apply to brand new rollators. Second-hand rollators supplied directly by Medicare Technology carry a six-month warranty period. If you are unsure whether your rollator is covered contact your dealer.
8. Under normal circumstances, no responsibility will be accepted where the rollator has required assistance as a direct result of:
   a. the rollator part not having been maintained in accordance with the manufacturers recommendations
   b. failure to use the manufacturer's specified parts
   c. the rollator or part having been damaged due to neglect, accident or improper use
   d. the rollator or part having been altered from the manufacturer's specification or repairs having been attempted before the dealer is notified

In the event of your rollator requiring attention, please contact your service agent / dealer and give all relevant details so they can act quickly.

The manufacturer reserves the right to alter without notice any weights, measurements or other technical data shown in this manual. All figures, measurements and capacities shown in this manual are approximate and do not constitute specifications.

The Active Rollator is distributed by:

Medicare Technology Limited

Armytage Road, Brighouse, West Yorkshire HD6 1PT
Tel +44 (0)1484 727301 Fax +44 (0)1484 727370
enquiries@medicaretechnology.com www.medicaretechnology.com

MINI / MIDI SCOOTERS
MAINSTREAM SCOOTERS
SCOOTER ACCESSORIES
WHEELCHAIRS
POWERCHAIRS
WHEELCHAIR ACCESSORIES
WALKING AIDS
RISE/RECLINE ARMCHAIRS
MASSAGE CHAIRS
HIGH SEAT FIRESIDE CHAIRS
FOOTWEAR
BATHROOM & TOILET AIDS
BEDROOM AIDS
DRUG FREE PAIN RELIEF